



# 检测报告

## Test Report

送检企业: 安徽大江医疗用品有限责任公司

样品名称: 棉 签

检验类别: 委托检验

厦门泓益检测有限公司

Xiamen Hongyi Testing Technology Co.Ltd

This report shall not be altered, increased or deleted. The results shown in this test report refer only to the sample(s) tested. Without written approval of Hongyi Testing, this test report shall not be copied except in full and published as advertisement Hongyi Physical & Chemical Lab.

厦门泓益检测有限公司

Xiamen Hongyi Testing Co., Ltd.

地址: 福建省厦门市同安区轻工食品工业区美禾路 99 号

ADDRESS: No. 99 Meihe Road, Light Industry Food Industrial Zone, Tongan District, Xiamen City, Fujian Province,

邮箱: hongyotest@123.com



报告编号: HYI20260608A1626

样品名称	: 棉签	规格型号	: 10S 单头竹棒
批号	: 260601	样品数量	: 2 包
委托单位	: 安徽大江医疗用品有限责任公司		
委托单位地址	: 安徽省滁州市定远县经开区东兴路 25 号		
生产单位	: 安徽大江医疗用品有限责任公司		
生产单位地址	: 安徽省滁州市定远县经开区东兴路 25 号		
检验类别	: 委托检验	测试日期	: 2026 年 06 月 02 日-2026 年 06 月 08 日
签发日期	: 2026 年 06 月 08 日		
检验依据	: GB 15979-2024 《一次性使用卫生用品卫生要求》		
检验项目	: 见后续		
样品特征和状态	: 完好		
环境温度	: 24.6℃	环境湿度	: 63.8%
测试结果	: 本次委托检验, 所检项目全部符合标准要求		

报告编制: 张华伟

报告审核: 杨泽群



This report shall not be altered, increased or deleted. The results shown in this test report refer only to the sample(s) tested. Without written approval of Hongyi Testing, this test report shall not be copied except in full and published as advertisement Hongyi Physical & Chemical Lab.



报告编号: HYI20260608A1626

检测结果:

序号	检测项目	标准要求	检测结果	单项判定
1	细菌菌落总数	≤200 cfu/g	76 cfu/g	符合
2	真菌菌落总数	≤100 cfu/g	28 cfu/g	符合
3	大肠菌群	不得检出	未检出	符合
4	铜绿假单胞菌 (Pseudomonas aeruginosa)	不得检出	未检出	符合
5	金黄色葡萄球菌 (Staphylococcus aureus)	不得检出	未检出	符合
6	溶血性链球菌 (Streptococcus hemolyticus)	不得检出	未检出	符合

## 样品照片



This report shall not be altered, increased or deleted. The results shown in this test report refer only to the sample(s) tested. Without written approval of Hongyi Testing, this test report shall not be copied except in full and published as advertisement Hongyi Physical & Chemical Lab.



## 注意事项:

1. 报告无“检验报告章”或检验单位公章无效。
2. 未经本实验室书面同意,不得部分复制本报告。
3. 报告未经主检、审核、批准人签字无效。
4. 报告涂改无效。
5. 对检测报告若有异议,应于收到报告之日起十五日内向检测单位提出,逾期不予受理。
6. 委托检测仅对来样负责。

地址:福建省厦门市同安区轻工食品工业区美禾路 99 号

Add: No. 99 Meihe Road, Light Industry Food Industrial Zone, Tongan District, Xiamen City, Fujian Province,

\*\*\*报告结束\*\*\*

This report shall not be altered, increased or deleted. The results shown in this test report refer only to the sample(s) tested. Without written approval of Hongyi Testing, this test report shall not be copied except in full and published as advertisement Hongyi Physical & Chemical Lab.

厦门泓益检测有限公司

Xiamen Hongyi Testing Co., Ltd.

地址:福建省厦门市同安区轻工食品工业区美禾路 99 号

ADDRESS: No. 99 Meihe Road, Light Industry Food Industrial Zone, Tongan District, Xiamen City, Fujian Province,

邮箱: hongyotest@123.com